



CARISOPRODOL AND MEPROBAMATE IN EQUINE URINE FOR GC/MS CONFIRMATIONS

Part Numbers:

XRDAH206 - 200 mg XtrackT[®] DAU Extraction Column in 6 mL cartridge

SPHPHO6001-10 - Select pH Buffer Pouches 100mM Phosphate pH 6.0

1. PREPARE SAMPLE:

To 2 mL of urine add internal standard(s) and 1 mL of 100 mM phosphate buffer (pH= 6)

Mix/vortex

Sample pH should be 6.0 ± 0.5

Adjust pH accordingly with 100 mM monobasic or dibasic sodium phosphate

Centrifuge at 3000 RPM for 10 minutes

2. CONDITION XTRACKT[®] DAU EXTRACTION COLUMN:

1 x 3 mL CH₃OH

1 x 3 mL D.I. H₂O

1 x 1 mL 100 mM phosphate buffer (pH= 6)

NOTE: Aspirate at full vacuum or pressure

3. APPLY SAMPLE:

Load at 1 to 2 mL/ minute

4. WASH COLUMN:

1 x 3 mL D.I. H₂O

1 x 1 mL 100 mM acetic acid

Dry column (5 minutes at full vacuum or pressure)

1 x 2 mL Hexane

5. ELUTE BARBITUATES:

1 x 3 mL Hexane/ Ethyl Acetate (50:50); Collect eluate at 1 to 2 mL / minute

6. DRY ELUATE:

Evaporate to dryness at $< 40^{\circ}\text{C}$


Reconstitute with 100 μL Ethyl Acetate

7. QUANTITATIVE:

Inject 1 to 2 μL onto gas chromatograph

CHROMATOGRAM

Carisoprodol, Meprobamate, and Hexobabital (Internal Standard)



Mass Spec Table

Compound	Primary Ion
Carisoprodol	221
Meprobamate	157
Hexobarbital	236